

# **CERTIFICATE OF ACCREDITATION**

### **The ANSI National Accreditation Board**

Hereby attests that

### Hexagon Manufacturing Intelligence, Inc. 46444 Hexagon Way Novi, MI 48377

(and the satellite as listed on the scope)

Fulfills the requirements of

## **ISO/IEC 17025:2017**

and the national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

### CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document. The current scope of accreditation can be verified at <u>www.anab.org</u>.





R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 14 December 2024 Certificate Number: AC-1745

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



#### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 AND ANSI/NCSL Z540-1-1994 (R2002)

#### Hexagon Manufacturing Intelligence, Inc.

46444 Hexagon Way Novi, MI 48377 Sarah White 248-449-9443

#### CALIBRATION

Valid to: December 14, 2024

Certificate Number: AC-1745

Length – Dimensional Metrology

Parameter / Equipment <sup>1</sup>	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method and/or Equipment
Length Bar Standard	Up to 1.2 m	(2.8 + 0.47 <i>L</i> ) μm	СММ
Articulated Arm CMM (AACMM) – Volumetric Performance	Up to 1.2 m	(4 + 0.67 <i>L</i> ) μm	ASME B89.4.22 Sections 5.3 and 5.4 Length Bar Standard
Articulated Arm CMM (AACMM) – Volumetric Performance	Up to 1.2 m	(2.6 + 0.63 <i>L</i> ) μm	ISO 10360-2 Sections 6.3 and 6.4 (adapted) Step Gage



www.anab.org



#### Length – Dimensional Metrology

Parameter / Equipment <sup>1</sup>	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method and/or Equipment
Articulated Arm CMM (AACMM) –	Sphere Diameter:		ISO 10360-12
Probing Size Error (P Size) <sup>2</sup>	Up to 51 mm	0. <mark>22 μm</mark>	Test Sphere
Probing Form Error (P Form) <sup>2</sup>	Up to 51 mm	0.13 μm	Test Sphere
Articulated Location Error (LDia) <sup>2</sup>	Up to 51 mm	0.16 µm	Test Sphere
Length Measurement Error Unidirectional (EUni)	Length: Up to 3 m	(2 + 2 <i>L</i> ) μm	Scale Bar with conical seat
Articulated Arm Coordinate Measuring Machines (AACMM) with Optical Distance Sensors:		un	ISO 10360-8 Annex D
Articulated Location Value (LDia) <sup>2</sup>	Sphere Diameter: 51 mm	0.25 μm	Test Sphere
Test Sphere			
Diameter <sup>2</sup>	25.4 mm 50.0 mm	1.6 μm 4.4 μm	СММ
Form <sup>2</sup>	25.4 mm 50.0 mm	1 μm 1 μm	
Length (Tracker with or without T-scan or T-probe)	(125 to 2 550) mm	14.1 μm	Scale Bar (Brunson kit, modular)
Length (Theodolites - Industrial Measurement Systems)	(125 to 2 550) mm	14.1 μm	Scale Bar (Brunson kit, modular)





#### Services performed at satellite location

9004 Research Drive Irvine, CA 92618 Sarah White 248-449-9443

#### CALIBRATION

#### Length – Dimensional Metrology

Parameter / Equipment <sup>1</sup>	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Length (Tracker with or without T-scan or T-probe)	(125 to 2 550) mm	14.1 μm	Scale Bar (Brunson kit, modular)
Length (Theodolites - Industrial Measurement Systems)	(125 to 2 550) mm	14.1 μm	Scale Bar (Brunson kit, modular)

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%. Notes:

- 1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
- 2. L = Length in meter.
- 3. Dimensions are nominal value
- 4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1745.



R. Douglas Leonard Jr., VP, PILR SBU



www.anab.org